

TEST REPORT

(Electronic version)



No: 200068526

VERIFICATION WEBSITE: www.gttc.net.cn

VERIFICATION CODE: SMYQ-8307-14

ISSUE DATE: 2020-04-10

扫码下载报告



APPLICANT: Sichuan Tenor Technology co.,Ltd
ADDRESS: No.50 Nanshan Road, Economic Development Zone, Mianyang City, Sichuan Province

SAMPLE INFORMATION:

FIFTY DISPOSABLE PROTECTIVE MASKS

BRAND: Tainuoer

COLOUR: BLUE

INGREDIENTS OF FIBRE: NON WOVEN FABRIC, MELTBLOWN

SIZE: 17.5cm×9.5cm (±0.5cm)

MANUFACTURE'S NAME: Sichuan Tenor Technology co.,Ltd

DATE RECEIVED/DATE TEST STARTED: 2020-04-06

CONCLUSION:

BACTERIAL FILTRATION EFFICIENCY M

PARTICLE FILTRATION EFFICIENCY M


NOTE: "M" -MEET THE STANDARD'S REQUIREMENT "F" -FAIL TO MEET THE STANDARD'S REQUIREMENT
"---" -NO COMMENT

REMARK:

THE DECISION INDICATORS ARE DERIVED FROM THE GROUP STANDARD REQUIRED BY CLIENT (T/CTCA 7-2019). OUR INSPECTION CAPACITY AUTHORIZED BY CMA COVERS THE INSPECTION ITEMS T/CTCA 7-2019.
THIS REPORT IS THE ENGLISH TRANSLATION VERSION OF THE REPORT 200068525.
ALL THE TESTED ITEMS ARE TESTED UNDER THE STANDARD CONDITION (EXCEPT FOR INDICATION).
COPIES OF THE REPORT ARE VALID ONLY RE-STAMPED.
THE EXPERIMENT WAS CARRIED OUT AT No. 1, ZHUJIANG ROAD, PANYU DISTRICT, GUANGZHOU, GUANGDONG, P. R. CHINA.

APPROVED BY:

WanLi Hu ENGINEER



PAGE 1 OF 3

TEST REPORT

(Electronic version)

No:200068526



TEST REPORT

(Electronic version)

No:200068526

BACTERIAL FILTRATION EFFICIENCY (%)

(YY 0469-2011 ANNEX B, TEST BACTERIA: STAPHYLOCOCCUS AUREUS ATCC 6538, TEST AREA: 40cm², FLOW RATE: 28.3L/min, MEAN PARTICLE SIZE: 3.0 μm, RESULT OF THE POSITIVE CONTROL: 1.9×10³ CFU, RESULT OF THE NEGATIVE CONTROL: <1CFU)

		REQUIREMENT
BFE ₁	98.8	≥95
BFE ₂	98.7	(T/CTCA 7-2019)
BFE ₃	98.5	

PARTICLE FILTRATION EFFICIENCY (%)

(YY 0469-2011 5.6.2, AIR FLOW: 30L/min, AEROSOL: NaCl, AEROSOL CONCENTRATION: 15mg/m³, TEMP: 23.2°C, RH: 36.2%)

		REQUIREMENT
MINIMUM	86.6	≥80
		(T/CTCA 7-2019)



—End of Report—

PAGE 3 OF 3